



Bijlage bij accreditatie-certificaat
Annexe au certificat d'accréditation
Annex to the accreditation certificate
Beilage zur Akkreditierungszertifikat

570-TEST

NBN EN ISO/IEC 17025:2005

Versie/Version/Fassung	2
Uitgivedatum / Date d'émission / Issue date / Ausgabedatum:	2017-03-10
Geldigheidsdatum / Date limite de validité / Validity date / Gültigkeitsdatum:	2019-03-16

Nicole Meurée-Vanlaethem

Voorzitster van het Accreditatiebureau
La Présidente du Bureau d'Accréditation
Chair of the Accreditation Board
Vorsitzende des Akkreditierungsbüro

**De accreditatie werd uitgereikt aan/ L'accréditation est délivrée à/
The accreditation is granted to/ Die akkreditierung wurde erteilt für:**

**METALogic NV
Wingepark, 43
3110 ROTSELAAR**

Secrétariat:

**Service public fédéral, Economie,
P.M.E., Classes moyennes et Energie**
Direction générale de la Qualité et de la Sécurité
Division Qualité et Innovation
Bd du Roi Albert II, 16 - 5^{ème} étage - B-1000 Bruxelles
Website: <http://economie.fgov.be>
Numéro d'entreprise: 0314.595.348

Accréditation BELAC Accreditation

Tél: +32 2 277 54 34
Fax: +32 2 277 54 41

Internet: <http://belac.fgov.be>
E-Mail: Belac@economie.fgov.be

Secretariaat:

**Federale Overheidsdienst, Economie,
K.M.O., Middenstand en Energie**
Algemene Directie Kwaliteit en Veiligheid
Afdeling Kwaliteit en Innovatie
Koning Albert II-laan 16 - 5^{de} verd. - B-1000 Brussel
Website: <http://economie.fgov.be>
Ondernemingsnummer: 0314.595.348

.be

Interne testcode	Monsters waarop de beproevingen worden uitgevoerd	Gemeten eigenschappen	Beschrijving beproevingsmethoden en meettechnieken
MP 4.5.1	Coated test panels Corrosion coupons	Mass loss of coupons exposed together with coated test panels	Exposure in a cyclic corrosion test according to SAE J2334 standard
MP 4.5.3 (CAT 1E2001) and MP 4.5.4 (determination of creepback values)	Coated test panels	Scribe creepback value	CAT 1E2001 §7 after exposure to SAE J2334 conditions